



6640 185th Ave NE, Redmond, WA 98052 T.425.895.8617, F.425.702.9358

ON-SITE CERTIFICATE #: OS2-00046031

CERTIFICATE OF CALIBRATION

STANDARD CALIBRATION

YAKIMA POLICE DEPARTMENT

200 SOUTH THIRD STREET YAKIMA, WA 98901

This certifies that the instrument listed herein was calibrated by Cascade Engineering Services' Calibration Laboratory, which is fully accredited in accordance with the recognized International Standards ISO/IEC 17025:2017 General Requirements for the Competence of Testing and Calibration Laboratories. Cascade Engineering Services' Calibration Laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and any additional program requirements in the field of calibration. Standards used to perform this calibration are certified by or traceable to NIST, natural physical constants, consensus standards or derived by the ratio type of calibrations. All calibrations are performed to manufacturer's specifications unless otherwise noted. Standard Calibration, while still traceable, does not meet all requirements for an Accredited Calibration per ISO/IEC 17025:2017, that is "As Found" data for equipment in tolerance and Measurement Uncertainties are not recorded. This certificate shall not be reproduced, except in full, without prior written approval of the laboratory.

DESCR	IPTION:		SPE	ED MEASURING DEVI	CE		
ASSET #	:	15-03		SERIAL NUMBER:	LF003216		
MANUFA	CTURER:	APPLIED CONCEPTS INC.		MODEL NUMBER:	XLR		
DEPARTMENT:		N/A		PATROL CAR #:	N/A		
ENVIRONMENT:		65.0 °F/39.0 %RH		BASIC ACCURACY:	REFERENCE CAL PROCEDURE		
CAL INTI		24 MONTHS		DUE DATE:_	June 03, 2023		
A + + h = = = = =			DELIVERED	ITOLEDANCE" as defined by	u the besie secure	ov stata	d about
LASER PULSE R	npletion of the cal SYSTEM OUT EFERENCE FRE	ibration, me PUTS QUENCY_	easured values were "IN 193.6870 Hz	I TOLERANCE", as defined by	y the basic accurac		d above. 3 μW
LASER PULSE R STAND	mpletion of the cal SYSTEM OUT EFERENCE FRE ARD(S) USED	ibration, me PUTS QUENCY_ FOR CE	easured values were "IN 193.6870 Hz RTIFICATION	OPTICAL POW	VER OUTPUT:		3 μW
LASER PULSE R	npletion of the cal SYSTEM OUT EFERENCE FRE	ibration, me PUTS QUENCY_ FOR CE	easured values were "IN 193.6870 Hz	OPTICAL POW			
LASER PULSE R STAND	npletion of the cal SYSTEM OUT EFERENCE FRE ARD(S) USED MODE	ibration, me PUTS QUENCY_ FOR CE	easured values were "IN 193.6870 Hz RTIFICATION MANUFACTURER	OPTICAL POW	VER OUTPUT:		3 μW DUE DATE
LASER PULSE R STAND I.D. MET1259 MET1260 MET1335	SYSTEM OUT EFERENCE FRE ARD(S) USED MODE NOVA-DISPLAY PD300-SH 7024707	ibration, me PUTS QUENCY_ FOR CEI L OF	193.6870 Hz RTIFICATION MANUFACTURER PHIR PHIR SER TECHNOLOGY INC	OPTICAL POW DESCR LASER POWER METER LASER POWER HEAD LASER SPEED MEASUREMENT	VER OUTPUT:		DUE DATE 03/17/2022
LASER PULSE R STAND I.D. MET1259 MET1260 MET1335 MET1336	SYSTEM OUT EFERENCE FRE ARD(S) USED MODE NOVA-DISPLAY PD300-SH 7024707 7005320	ibration, me PUTS QUENCY_ FOR CE	easured values were "IN 193.6870 Hz RTIFICATION MANUFACTURER PHIR PHIR SER TECHNOLOGY INC SER TECHNOLOGY INC	OPTICAL POW DESCR LASER POWER METER LASER POWER HEAD	VER OUTPUT:		DUE DATE 03/17/2022 03/22/2022
LASER PULSE R STAND I.D. MET1259 MET1260 MET1335 MET1336	SYSTEM OUT EFERENCE FRE ARD(S) USED MODE NOVA-DISPLAY PD300-SH 7024707 7005320	ibration, me PUTS QUENCY_ FOR CE	193.6870 Hz RTIFICATION MANUFACTURER PHIR PHIR SER TECHNOLOGY INC	OPTICAL POW DESCR LASER POWER METER LASER POWER HEAD LASER SPEED MEASUREMENT	VER OUTPUT:		DUE DATE 03/17/2022 03/22/2022
LASER PULSE R STAND I.D. MET1259 MET1260 MET1335 MET1336 PROCE	SYSTEM OUT EFERENCE FRE ARD(S) USED MODE NOVA-DISPLAY PD300-SH 7024707 7005320	ibration, me PUTS QUENCY_ FOR CE	193.6870 Hz RTIFICATION MANUFACTURER PHIR PHIR SER TECHNOLOGY INC SER TECHNOLOGY INC ERTIFICATION	OPTICAL POW DESCR LASER POWER METER LASER POWER HEAD LASER SPEED MEASUREMENT	VER OUTPUT:	428	DUE DATE 03/17/2022 03/22/2022

THIS LABORATORY IS A2LA ACCREDITED TO ISO/IEC 17025:2017 (GENERAL REQUIREMENTS FOR THE COMPETENCE OF TESTING AND CALIBRATION LABORATORIES), CERTIFICATE #: 2560.01

LOCATION:

CALIBRATION DATE:

I certify (or declare) under penalty of perjury under the laws of the State of Washington that the above information is true and correct

METROLOGIST: JOHN GRAY

F540-1.3/REVISION M

YAKIMA, WA

Thursday, June 03, 2021